

# Significance of Use

Cyanuric acid (CYA) is best known as a stabilizing reagent for chlorine. It is widely applied in swimming pool and spa treatment programs to slow down the decomposition of hypochlorous acid. In outside pool areas, this process is accelerated by the effects of UV rays. When applied properly it can save up to 80% of normal chlorine consumption in pools during peak months.

Cyanuric acid is also used in chlorinated beaches, selective herbicides and whitening agents.

#### Specifications HI96722 Cyanuric Acid 0 to 80 mg/L (ppm) Range Resolution 1 mg/L (ppm) Accuracy @ 25°C (77°F) ±1 mg/L ±15% of reading Light Source tungsten lamp Light Detector silicon photocell with narrow band interference filter @ 525 nm Power Supply 9V battery after ten minutes of non-use in measurement mode; after one hour Auto-off of non-use in calibration mode; with last reading reminder Environment 0 to 50°C (32 to 122°F); RH max 95% non-condensing Dimensions 193 x 104 x 69 mm (7.6 x 4.1 x 2.7") Weight 360 g (12.7 oz.) Method adaptation of the turbidimetric method HI96722 is supplied with sample cuvettes (2) with caps, 9V battery, instrument quality Ordering certificate and instruction manual. Information CAL Check™ standards and testing reagents sold separately HI96722-11 CAL Check™ standard cuvettes Reagents and HI93722-01 reagents for 100 tests Standards HI93722-03 reagents for 300 tests

## HI96722

# Cyanuric Acid Photometer

#### CAL Check

 Allows for performance verification and calibration of the meter using NIST traceable standards.

#### GLP

· Review of the last calibration date.

#### · Auto-shut off

 Automatic shut off after 10 minutes of non-use when the meter is in measurement mode. Prevents wastage of batteries in the event the meter is accidentally left on.

# • Battery status indicator

 Indicates the amount of battery life left.

#### · Built-in timer

 Display of time remaining before a measurement is taken. Ensures that all readings are taken at the appropriate reaction intervals for the test being performed.

## Error messages

 Messages on display alerting to problems including no cap, high zero, and standard too low.

# · Cooling lamp indicator

 To maintain the desirable wavelength to be used for absorbance, it is necessary to ensure components are not overheated from the heat generated by the tungsten lamp. Each photometer is designed to allow a minimal amount of time for components to cool. The cooling lamp indicator is displayed prior to a reading being taken.

# • Units of measure

 Appropriate unit of measure is displayed along with reading.

The HI96722 portable photometer is for the measurement of cyanuric. Hanna's portable photometers feature an advanced optical system; the combination of a special tungsten lamp, a narrow band interference filter, and silicon photodetector ensure accurate photometric readings every time. The Hanna exclusive CAL Check™ feature utilizes readymade, NIST traceable standards to verify both meter validation and calibration. The exclusive cuvette locking system ensures that the cuvette is inserted into the measurement cell in the same position every time to maintain a consistent path length.

ติดต่อบริษัท นีโอนิคส์ จำกัด

Tel: 02-077-7602 หรือ 061-8268939

E-mail: sale@neonics.co.th เว็บไซต์ www.neonics.biz

